

Complement C3

For the Quantitative Determination of Human Complement C3 in Serum

Cat. No. KAI-009

INTENDED USE

For the quantitative determination of human complement C3 (3rd complement component) in serum by immunoturbidimetric assay. Complement is a group of serum proteins that destroy infectious agents. Measurement of these proteins aids in the diagnosis of immunological disorders, especially those associated with deficiencies of complement components. FOR *IN VITRO* DIAGNOSTIC USE.

INTRODUCTION AND SUMMARY

C3 is the third complement component. It is one of a group of serum proteins that are active in the body's immune response. Complement C3 is an opsonic protein, playing a role in destroying infectious agents. The level of the 3rd complement component (C3) in serum can be used to help identify immunological disorders, especially those associated with deficiencies of complement components.^{1,2,3}

Complement C3 has been measured using a variety of methods, including radioimmunoassay (RIA), radial diffusion, nephelometric, and enzyme-linked immunosorbent assay.^{2,3} The **K-ASSAY**® Complement C3 assay uses an immunoturbidimetric assay format.

PRINCIPLE OF TEST

The **K-ASSAY**® Complement C3 assay quantifies the 3rd complement component in the patient's serum based by immunoturbidimetric assay. Calibrators, controls, and patient samples are pipetted into sample cups. Microvolumes of samples and antibody diluent are automatically pipetted into individual cuvettes. Following an initial incubation and measurement of sample blank, antiserum is added to the cuvettes. The samples (antigen) and antiserum are then mixed in the reaction cuvettes. Insoluble antigen-antibody (immune) complexes form. The immune complexes cause an increase in light scattering that correlates with the concentration of serum complement C3. Following an incubation period lasting approximately 5 min., the absorbance of the solution is measured at 600 nm.

A calibration curve is generated by assaying a series of calibrators with known concentrations of proteins and using the instrument's data reduction capability or manually plotting the change in absorbance versus concentration. Concentration of the control and patient samples are interpolated from the calibration curve. The antiserum used in the kit is a goat polyclonal antibody specific for human complement C3.

The **K-ASSAY**® Complement C3 assay can be run using a two-reagent clinical chemistry analyzer. Six calibrators are provided in the **K-ASSAY**® Multi-Analyte Calibrator. These calibrators are used to prepare a calibration curve for quantifying the levels of complement C3 present in the patient's serum sample.

KIT COMPOSITION

Reagents (Liquid Stable)

R1: Buffer Reagent 4 x 20 mL
Tris(hydroxymethyl)aminomethane

R2: Antiserum Reagent 2 x 10 mL
Anti-human complement C3 goat antiserum (35%)

WARNINGS AND PRECAUTIONS

FOR *IN VITRO* DIAGNOSTIC USE. Rx only.

Not to be used internally in humans or animals. Normal precautions exercised in handling laboratory reagents should be followed.

Do not mix or use reagents from one test kit with those from a different lot number.

Do not use reagents past their expiration date stated on each reagent container label.

Do not pipette by mouth. Avoid ingestion and contact with skin.

Reagents in this kit contain sodium azide as a preservative. Sodium azide may form explosive compounds in metal drain lines. When disposing of reagents through plumbing fixtures, flush with copious amounts of water.

For further information, refer to "Decontamination of Laboratory Sink Drains to Remove Azide Salts," in the Manual Guide-Safety Management No. CDC-22 issued by the Centers for Disease Control, Atlanta, Georgia.

REAGENT PREPARATION

Reagents are ready to use and do not require reconstitution.

STORAGE AND HANDLING

All reagents should be stored refrigerated (2-8°C). Return all reagents to 2-8°C promptly after use. Unopened reagents can be used for up to 18 months from the date of manufacture, as indicated by the expiration date on package and bottle labels.

REAGENT STABILITY

Opened reagents can be used for one month if stored at 2-8°C. Discard reagents if they become contaminated. Evidence of cloudiness or particulate material in solution is cause to discard.

INSTRUMENT

Measurement of absorbance is to be made with an instrument able to accurately read absorbance at 600 nm. Refer to the instrument manual from the manufacturer regarding the following:

- Use or function
- Installation procedures and requirements
- Principles of operation
- Performance characteristics, operating instructions
- Calibration procedures including materials and / or equipment to be used
- Operational precautions, limitations, and hazards
- Service and maintenance information

SPECIMEN COLLECTION AND PREPARATION

It is recommended that specimen collection be carried out in accordance with NCCLS document M29-T2. No method can offer complete assurance that human blood samples will not transmit infection. Therefore, all blood samples should be considered potentially infectious.

Serum is required for this assay. Soon after the blood is drawn, it should be allowed to clot, centrifuged, and the serum separated from the clot to a plastic tube (not glass) within 2 hours. Serum should be stored refrigerated (2-8°C) and can be used within one week or should be stored frozen for up to 2 months.

Use plastic tubes for storing the sample; do not use glass.

AUTOMATED ANALYZER APPLICATION

Suitable for two-reagent automated analyzers that use a multi-point calibration method.

PROCEDURE

Materials Supplied

Reagent 1 (R-1) Buffer Reagent 4 x 20 mL
Reagent 2 (R-2) Antiserum Reagent 2 x 10 mL

Materials Required But Not Supplied

Calibrators: **K-ASSAY**® Multi-Analyte Calibrator, Cat. No. KAI-016C (6 calibrators containing human serum with known levels of complement C3).

Two-Reagent Clinical Chemistry Analyzer:

- Capable of accurate absorbance readings at 600 nm
- Capable of accurately dispensing the required volumes
- Capable of maintaining 37°C

Assay Procedure

Note: Allow all reagents and specimens to warm to room temperature. Mix all reagents gently before using.

An example of automated application (Hitachi 717):

Sample	5 µL
↓	
• ← R1 (Buffer Reagent)	250 µL
↓	37°C, 5 min.
• ← R2 (Antiserum Reagent)	70 µL
↓	37°C, 5 min.
2-point endpoint, 600 nm	

Automated Method (Example)

Chemistry Parameters for Automatic Analyzer

INSTRUMENT	Hitachi 717
TEMPERATURE	37°C
TEST	(C3)
ASSAY CODE	(2 POINT) : (24) - (50)
SAMPLE VOLUME	(5) ()
R1 VOLUME	(250) () (NO)
R2 VOLUME	(70) () (NO)
WAVELENGTH	() (600)
CALIB. METHOD	(NONLINEAR) (1) (6)
STD.(1) Conc.-POS.	(*1) - (1)
STD.(2) Conc.-POS.	(*2) - (2)
STD.(3) Conc.-POS.	(*3) - (3)
STD.(4) Conc.-POS.	(*4) - (4)
STD.(5) Conc.-POS.	(*5) - (5)
STD.(6) Conc.-POS.	(*6) - (6)
SD LIMIT	(999)
DUPLICATE LIMIT	(10000)
SENSITIVITY LIMIT	(0)
ABS. LIMIT (SLOPE)	(32000) (INCREASE)
PROZONE LIMIT	(-32000) (LOWER)
EXPECTED VALUE	(-99999) (99999)
PANIC VALUE	(-99999) (99999)
INSTRUMENT FACTOR	(1.00)

* 1-6 Input concentration of calibrators

Parameters for other automated analyzers are available.

CALIBRATION

It is recommended that complement C3 levels be determined using a multi-point calibration curve prepared using the **K-ASSAY**® Multi-Analyte Calibrator. It is recommended that the user determine calibration curve frequency as this depends on the instrument and type/number of other assays being performed. Initially, calibration should be performed each day.

QUALITY CONTROL

Normal and abnormal controls of known concentration should be included in each assay performed. These controls should fall within the stated values assigned to the controls. The validity of the assay is in question if the value for the controls generated by the assay's calibration curve does not fall within the stated range. Recalibrate if the value determined for the controls falls outside the stated recovery range.

RESULTS / CALCULATIONS

Complement C3 levels are determined using the prepared calibration curve.

LIMITATIONS OF PROCEDURE

The measurable range for complement C3 is between 30 and 350 mg/dL. Grossly lipemic samples and samples with very high triglyceride concentrations should be diluted 1 part sample with 1 part isotonic saline or filtered to decrease nonspecific light scattering. Multiply results by 2 to compensate for the dilution.

If the complement C3 concentration of a patient sample is greater than the highest calibrator value, dilute 1 part sample with 4 parts isotonic saline and reassay. Multiply results by 5 to compensate for the dilution.

The absorbance of isotonic saline should be below 0.05 and the absorbance of the 208 mg/dL calibrator should be between 0.25 and 0.40 after allowing for the reagent blank.

PERFORMANCE

Precision

The precision for the **K-ASSAY**® Complement C3 assay was determined using packaged reagents, pooled human serum, and a Hitachi 717 analyzer.

Precision Assay: Within Run

Sample I	Sample II	Sample III
N = 20	N = 20	N = 20
Mean = 33.7	Mean = 75.6	Mean = 112.4
SD = 1.12	SD = 1.68	SD = 2.18
CV = 3.33%	CV = 2.22%	CV = 1.94%

Precision Assay: Between Runs

Sample I	Sample II	Sample III
N = 10	N = 10	N = 10
Mean = 35.4	Mean = 84.3	Mean = 114.4
SD = 1.29	SD = 1.95	SD = 1.70
CV = 3.63%	CV = 2.31%	CV = 1.48%

Accuracy / Correlation

A comparison of the **K-ASSAY**® Complement C3 assay and an INCSTAR Complement C3 Test Kit was performed using a Hitachi 717. The test results provided the following data:

$$y = 0.842x + 4.727$$
$$r = 0.925$$
$$n = 60$$

x = INCSTAR Complement C3 Test Kit
y = **K-ASSAY**® Complement C3 Assay

x min = 78	y min = 74
max = 188	max = 171
mean = 135	mean = 118

Assay Range

30 - 350 mg/dL

INTERFERENCE

Bilirubin F and C:	No interference up to 20 mg/dL
Hemoglobin:	No interference up to 500 mg/dL
Lipemia:	No interference up to 5%

EXPECTED VALUES

The expected value as reported is between 74 and 148 mg/dL. Each laboratory should establish its own expected values using this kit.

REFERENCES

1. Ross, G.D. "Complement and Complement Receptors." *Curr. Opin. Immunol.* 2: 50-62, 1989.
2. Herbert, A., Cosio, F.G, and J.C. Negg. "Diagnostic Significance of Hypocomplementemia." *Kidney Int.* 39: 811-21, 1991.
3. Fijen, CAP *et al.* "Complement Deficiencies in Patients over Ten Years Old with Meningococcal Disease Due to Uncommon Serogroup." *Lancet* 2: 585-8, 1989.
4. Bergstrom, K., *et al.* "An Automated Turbidimetric Immunoassay for Plasma Proteins." *Scand. J. Clin. Lab. Invest.*, 40:637, 1980.
5. Finley, P. *et al.* *Clin Chem.* 22: 1037, 1976.
6. Heidelberger, M. and F. Kendall. *J. Exp. Med.*, 61: 563, 1935.
7. Killingsworth, L.M. and Savory, J., "Nephelometric Studies on the Precipitin Reactions," *J. Clin. Chem.*, 19: 403-407, 1973.
8. Sternberg, J.C. "A Rate Nephelometer for Measuring Specific Proteins by Immunoprecipitin Reaction," *Clin. Chem.*, 23:1456-64, 1977.

LABELING SYMBOLS

	Lot Number
	Reagent
	Expiration or "Use By" Date
	Catalog Number
	For <i>In Vitro</i> Diagnostic Use
	Temperature Limitation. Store between 2 and 8 degrees C
	Potential Human Biohazard
	Manufacturer
	Consult Package Insert for Instructions for Use
	Authorized Representative in the European Community

EU AUTHORIZED REPRESENTATIVE



ECIREP

Advena Ltd.

Tower Business Centre, 2nd Flr.,
Tower Street, Swatar, BKR 4013 Malta

ORDERING / PRICING / TECHNICAL INFORMATION



KAMIYA BIOMEDICAL COMPANY

12779 Gateway Drive
Seattle, WA 98168 USA
TEL: (206) 575-8068 / (800) 526-4925
FAX: (206) 575-8094