

IgG

For the Quantitative Determination of Human IgG in Serum

Cat. No. KAI-014

INTENDED USE

For the quantitative determination of human IgG in serum by immunoturbidimetric assay. Measurement of IgG aids in the diagnosis of abnormal protein metabolism and the body's lack of ability to resist infectious agents. FOR *IN VITRO* DIAGNOSTIC USE.

INTRODUCTION AND SUMMARY

Immunoglobulins are an important part of the body's immune response. The immunoglobulin G (IgG) molecule is composed of two light chains (kappa and lambda) and two gamma heavy chains. Approximately 80% of the serum immunoglobulin is IgG. The IgG's main function is to serve as a defense against microorganisms, direct neutralization of toxins, and induce the complement fixation pathway.

IgG levels in serum may be quantitated using a variety of methods such as turbidimetric, nephelometric, immunodiffusion or immunoassay.^{1,2,3,4} This assay uses an immunoturbidimetric method, taking advantage of the light scattering properties of antigen-antibody complexes.⁵ Antibody will bind specifically to the antigen in question, forming a complex. This complex can be quantitated by measuring light absorbance at 700 nm. The sensitivity and the rate of forming the immune-complex can be increased by the addition of the polymer, polyethylene glycol (PEG).⁶

PRINCIPLE OF TEST

Human serum, containing IgG, is diluted with buffer containing polyethylene glycol (PEG) and mixed with specific polyclonal goat anti-IgG antiserum. The antigen (IgG) and the specific goat antibody form complexes. The formation of the complexes is accelerated and enhanced by PEG. This allows for the reaction to rapidly reach its endpoint with greater sensitivity and less concern for false negative values due to antigen excess. The immune complexes cause an increase in light scattering that correlates with the concentration of IgG in the serum. Light scattering is measured by reading turbidity at 700 nm. The kit can be run using an analyzer. Six calibrators are provided in the **K-ASSAY®** Multi-Analyte Calibrator. These calibrators are to be used to prepare a calibration curve for quantifying the levels of IgG present in the patient serum sample.

KIT COMPOSITION

Reagents (Liquid stable)

R1: Buffer Reagent, pH 7.6 100 mM Tris(hydroxymethyl)aminomethane	4 x 20 mL
R2: Antiserum Reagent, pH 7.6 20% anti-human IgG goat antiserum	4 x 20 mL

WARNINGS AND PRECAUTIONS

FOR *IN VITRO* DIAGNOSTIC USE. Rx only.

Not to be used internally in humans or animals. Normal precautions exercised in handling laboratory reagents should be followed.

Do not mix or use reagents from one test kit with those from a different lot number.

Do not use reagents past their expiration date stated on each reagent container label.

Do not pipette by mouth. Avoid ingestion and contact with skin.

Reagents in this kit contain sodium azide as a preservative. Sodium azide may form explosive compounds in metal drain lines. When disposing of reagents through plumbing fixtures, flush with copious amounts of water.

REAGENT PREPARATION

Reagents are ready to use and do not require reconstitution.

STORAGE AND HANDLING

All reagents should be stored refrigerated (2-8°C). Return all reagents to 2-8°C promptly after use. Unopened reagents can be used for up to 18 months from the date of manufacture, as indicated on the expiration date on the package and bottle labels.

REAGENT STABILITY

Opened reagents can be used for 1 month if stored at 2-8°C. Discard reagents if they become contaminated. Evidence of cloudiness or particulate material in solution is cause to discard.

SPECIMEN COLLECTION AND PREPARATION

After drawing blood, allow it to completely coagulate. Centrifuge the coagulated blood and collect the supernatant. The supernatant can be directly used for testing without dilution. Samples may be stored for up to one week refrigerated. Serum samples stored for extended periods should be frozen at -20°C.

For storage of samples for more than a few days, use of plastic tubes is recommended instead of glass.

AUTOMATED ANALYZER APPLICATION

Suitable for two-reagent automated analyzers that use a six-point calibration method. Measurements of absorbance are to be made with a spectrophotometer able to accurately read absorbance at 700 nm. Refer to the instrument manual from the manufacturer regarding the following:

- Use or function
- Installation procedures and requirements
- Principles of operation
- Performance characteristics, operating instructions
- Calibration procedures including materials and / or equipment to be used
- Operational precautions, limitations, and hazards
- Service and maintenance information

PROCEDURE

Materials Supplied

Reagent 1 (R-1) Buffer Reagent	4 x 20 mL
Reagent 2 (R-2) Antiserum Reagent	4 x 20 mL

Materials Required But Not Supplied

Calibrators: **K-ASSAY®** Multi-Analyte Calibrator, Cat. No. KAI-016C, 6 Calibrators; Approximate values: 0, 450, 1,200, 2,000, 3,000, 4,000 mg/dL (For actual values see Package Insert)

Two Reagent Clinical Chemistry Analyzer:

- Capable of accurate absorbance readings at 700 nm
- Capable of accurately dispensing the required volumes
- Capable of maintaining 37°C

Test Tubes: plastic or glass (for short term storage only)

Assay Procedure

An example automated application (Hitachi 717):

Sample	4 µL
↓	
• ← R1 (Buffer Reagent)	250 µL
↓	37°C, 5 min.
• ← R2 (Antiserum Reagent)	250 µL
↓	37°C, 5 min.
	2-point endpoint, 700 nm

Note: Allow all reagents and specimens to come to room temperature. Mix all reagents gently before using.

Automated Method (Example)

Chemistry Parameters for Automatic Analyzer

INSTRUMENT	Hitachi 717
TEMPERATURE	37°C
TEST	(IgG)
ASSAY CODE	(2 POINT) : (24) - (50)
SAMPLE VOLUME	(4) ()
R1 VOLUME	(250) () (NO)
R2 VOLUME	(250) () (NO)
WAVELENGTH	() (700)
CALIB. METHOD	(NONLINEAR) (1) (6)
STD.(1) Conc.-POS.	(*1) - (1)
STD.(2) Conc.-POS.	(*2) - (2)
STD.(3) Conc.-POS.	(*3) - (3)
STD.(4) Conc.-POS.	(*4) - (4)
STD.(5) Conc.-POS.	(*5) - (5)
STD.(6) Conc.-POS.	(*6) - (6)
SD LIMIT	(999)
DUPLICATE LIMIT	(10000)
SENSITIVITY LIMIT	(0)
ABS. LIMIT (SLOPE)	(32000) (INCREASE)
PROZONE LIMIT	(-32000) (LOWER)
EXPECTED VALUE	(-99999) (99999)
PANIC VALUE	(-99999) (99999)
INSTRUMENT FACTOR	(1.00)

* 1-6 Input concentration of calibrators

Parameters for other automated analyzers are available.

CALIBRATION

It is recommended that a multi-point calibration curve be made using the **K-ASSAY®** Multi-Analyte Calibrator. It is recommended that the user determine calibration frequency as this depends on the instrument and type/number of other assays being performed. Initially, calibration should be done each day.

QUALITY CONTROL

A quality control program is recommended for all clinical testing and laboratories. It is recommended that control serums, both normal and abnormal, be run with each batch of samples to monitor the procedure.

The values obtained for controls should fall within the manufacturer's specified range. A laboratory may establish its own control serum by assaying the serum a sufficient number of times to generate a valid mean and acceptable range.

RESULTS / CALCULATIONS

IgG levels are determined using the prepared calibration curve.

LIMITATIONS OF PROCEDURE

The measurable range for this IgG test kit is between 100 mg/dL and 4,500 mg/dL. If IgG concentrations are greater than the highest calibrator value, dilute 1 part sample with 4 parts isotonic saline and re-assay. Multiply the result by 5 to compensate for the dilution.

PERFORMANCE

Precision Assay: (Within Run)

SERUM I	SERUM II	SERUM III
N = 20	N = 20	N = 20
Mean = 901	Mean = 1,883	Mean = 2,927
SD = 16.9	SD = 30.0	SD = 78.1
CV = 1.9%	CV = 1.6%	CV = 2.7%

Precision Assay: (Between Runs)

SERUM I	SERUM II	SERUM III
N = 7	N = 7	N = 7
Mean = 437	Mean = 1,500	Mean = 2,002
SD = 13.4	SD = 20.9	SD = 28.2
CV = 3.1%	CV = 1.4%	CV = 1.4%

Accuracy / Correlation

A comparison of this **K-ASSAY**® IgG Assay and an INCSTAR IgG Test Kit was performed on a Hitachi 704 automated analyzer and a COBAS Mira. The test results provided the following data:

$y = 1.0789x - 88.01$
 $r = 0.99322$
 $n = 46$
 $x = \text{INCSTAR IgG Test Kit}$
 $y = \text{K-ASSAY}^{\circledR} \text{ IgG Assay}$

$x \text{ min} = 476$	$y \text{ min} = 500$
$\text{max} = 3,957$	$\text{max} = 4,507$
$\text{mean} = 1,542$	$\text{mean} = 1,576$

Assay Range

100 - 4,500 mg/dL

INTERFERENCE

Bilirubin:	No interference up to 20 mg/dL
Hemoglobin:	No interference up to 500 mg/dL
Lipemia:	No interference up to 5%











EXPECTED VALUES

The expected value is between 806 mg/dL and 1,813 mg/dL. These values were determined using normal serum from 89 healthy donors. Each laboratory should establish its own expected values using this kit.

REFERENCES

1. Killingsworth, L.M. and J. Savory. *J. Clin. Chem.* 19:403-407, 1973.
2. Sternberg, J.C. *Clin. Chem.* 23:1456, 1977.
3. Bergstrom, K., et al. *Scand. J. Clin. Lab. Invest.*, 40:673, 1980.
4. Malkus, H., et al. *Clinica Chimica Acta*, 88:523-530, 1978.
5. Killingsworth, L.M. and Savory, J. *Clin. Chem.* 20:1181, 1974.
6. Lizana, J. and K. Helling. *Clin. Chem* 20:1181, 1974.

LABELING SYMBOLS

	Lot Number
	Reagent
	Expiration or "Use By" Date
	Catalog Number
	For <i>In Vitro</i> Diagnostic Use
	Temperature Limitation. Store between 2 and 8 degrees C
	Potential Human Biohazard
	Manufacturer
	Consult Package Insert for Instructions for Use
	Authorized Representative in the European Community

EU AUTHORIZED REPRESENTATIVE



Advena Ltd.
Tower Business Centre, 2nd Flr.,
Tower Street, Swatar, BKR 4013 Malta

ORDERING / PRICING / TECHNICAL INFORMATION



KAMIYA BIOMEDICAL COMPANY
12779 Gateway Drive
Seattle, WA 98168 USA
TEL: (206) 575-8068 / (800) 526-4925
FAX: (206) 575-8094