



## CALIBRATOR VALUES











The assigned values for the **K-ASSAY**® D-Dimer Calibrator is provided in D-Dimer Units (DDU).

### Lot 640K

Calibrator 1	0.00 µg/mL DDU
Calibrator 2	2.15 µg/mL DDU
Calibrator 3	4.13 µg/mL DDU
Calibrator 4	7.22 µg/mL DDU
Calibrator 5	14.81 µg/mL DDU
Calibrator 6	30.59 µg/mL DDU

The expected values for the **K-ASSAY**® D-Dimer Calibrator are continually being revised through ongoing quality assurance. As a result, the expected values may change from lot to lot. Please refer to the package insert for each lot for the exact calibrator values.

## LABELING SYMBOLS

	Lot Number
	Calibrator
	Expiration or "Use By" Date
	Catalog Number
	For <i>In Vitro</i> Diagnostic Use
	2-8°C Temperature Limitation. Store between 2 and 8 degrees C
	Potential Human Biohazard
	Manufacturer
	Consult Package Insert for Instructions for Use
	Authorized Representative in the European Community

## EU AUTHORIZED REPRESENTATIVE



### Advena Ltd.

Tower Business Centre, 2<sup>nd</sup> Flr.,  
Tower Street, Swatar, BKR 4013 Malta

## ORDERING / PRICING / TECHNICAL INFORMATION



### KAMIYA BIOMEDICAL COMPANY

12779 Gateway Drive  
Seattle, WA 98168 USA  
TEL: (206) 575-8068 / (800) 526-4925  
FAX: (206) 575-8094

Printed February 2017