



KAMIYA BIOMEDICAL COMPANY

Cat Alpha-1 Acid Glycoprotein ELISA

**For the quantitative determination of alpha-1 acid glycoprotein
in cat serum or plasma.**

Cat. No. KT-700

For Research Use Only.

PRODUCT INFORMATION

Cat Alpha-1 Acid Glycoprotein ELISA **Cat. No. KT-700**

PRODUCT

The **K-ASSAY®** Cat Alpha-1 Acid Glycoprotein ELISA is an enzyme immunoassay for the quantitative determination of alpha-1 acid glycoprotein (α -1-AGP) in cat serum or plasma. For research use only.

INTRODUCTION

α -1-AGP is an acute phase protein that is elevated in cat serum as a result of injury, infection or disease. A 19-fold increase of α -1-AGP was reported in serum from hospitalized cats and a 5.7 fold increase after injection of normal cats with lipopolysaccharide. Of the positive acute phase proteins investigated (SAA, α -1-AGP, haptoglobin and CRP), they reported that α -1-AGP was the most responsive. More recently, α -1-AGP was reported to be a powerful discriminating biomarker for studying feline infectious peritonitis.

PRINCIPLE

The **K-ASSAY®** Cat α -1-AGP ELISA uses affinity purified anti-cat α -1-AGP antibodies for solid phase (microtiter wells) immobilization and horseradish peroxidase (HRP) conjugated anti-cat α -1-AGP antibodies for detection. The test sample is diluted and incubated in the microtiter wells for 45 minutes. The microtiter wells are subsequently washed and HRP conjugate is added and incubated for 45 minutes. This results in α -1-AGP molecules being sandwiched between the immobilization and detection antibodies. The wells are then washed to remove unbound HRP-labeled antibodies and TMB Reagent is added and incubated for 20 minutes at room temperature. This results in the development of a blue color. Color development is stopped by the addition of Stop Solution, changing the color to yellow, and optical density is measured spectrophotometrically at 450 nm. The concentration of α -1-AGP is proportional to the optical density of the test sample.

COMPONENTS

- Anti-cat α -1-AGP antibody coated microtiter plate with 96 wells (provided as 12 detachable strips of 8)
- Enzyme Conjugate Reagent, 11 mL
- Cat α -1-AGP Calibrator (lyophilized)
- Diluent (10X), 25 mL
- Wash Solution (20X), 50 mL
- TMB Reagent (One-Step), 11 mL
- Stop Solution (1N HCl), 11 mL

MATERIALS REQUIRED BUT NOT PROVIDED

- Precision pipettes and tips
- Distilled or de-ionized water
- Polypropylene or glass tubes
- Vortex mixer
- Absorbent paper or paper towels
- Micro-Plate incubator/shaker mixing speed of ~150 rpm
- A microtiter plate reader capable of measuring absorbance at 450 nm, with a bandwidth of 10 nm or less and an OD range of 0-4 OD
- Graph paper (PC graphing software is optional)

GENERAL INSTRUCTIONS

All reagents should be allowed to reach room temperature (18-25°C) before use.

WASH SOLUTION PREPARATION

The wash solution is provided as a 20X stock. Prior to use dilute the contents of the bottle (50 mL) with 950 mL of distilled or de-ionized water. The diluted wash solution is stable for 1 day at room temperature.

DILUENT PREPARATION

The diluent is provided as a 10X stock. Prior to use estimate the final volume of diluent required for your assay and dilute one volume of the 10X stock with nine volumes of distilled or de-ionized water.

CALIBRATOR PREPARATION

1. The Cat α -1-AGP calibrator is provided as a lyophilized stock. Add the volume of distilled or de-ionized water indicated on the vial label and mix gently until dissolved.
2. Label 6 polypropylene or glass tubes as 50, 25, 12.5, 6.25, 3.13 and 1.56 ng/mL.
3. Into the tube labeled 50 ng/mL, prepare the 50 ng/mL working calibrator as detailed on the stock vial label.
4. Dispense 250 μ L of 1X diluent into the remaining tubes.
5. Prepare the 25 ng/mL calibrator by diluting and mixing 250 μ L of the 50 ng/mL calibrator with 250 μ L of diluent in the tube labeled 25 ng/mL.
6. Similarly prepare the 12.5, 6.25, 3.13 and 1.56 ng/mL calibrators by serial dilution.

Please Note: The reconstituted calibrator remains stable for at least 7 days at 4°C but should be aliquoted and stored frozen at -20°C if future use is intended.

SAMPLE PREPARATION

General Note: Serum concentrations of α -1-AGP may range from 0.1 to 5 mg/mL or greater. In order to obtain values within the range of the calibration curve we suggest that samples initially be diluted 100,000 fold using the following procedure for each sample to be tested:

1. Dispense 998 μ L and 497.5 μ L of 1X diluent into separate polypropylene or glass tubes.
2. Pipette and mix 2 μ L of the serum/plasma sample into the tube containing 998 μ L of diluent. This provides a 500 fold diluted sample.
3. Mix 2.5 μ L of the 500 fold diluted sample with the 497.5 μ L of diluent in the second tube. This provides a 100,000 fold dilution of the sample.
4. Repeat this procedure for each sample to be tested.

ASSAY PROCEDURE

1. Secure the desired number of coated wells in the holder.
2. Dispense 100 μ L of calibrators and samples into the wells (we recommend that samples be tested in duplicate).
3. Incubate on an orbital micro-plate shaker at 100-150 rpm at room temperature (18-25°C) for 45 minutes.
4. Aspirate the contents of the microtiter wells and wash the wells 5 times with 1X wash solution using a plate washer (400 μ L/well). The entire wash procedure should be performed as quickly as possible.
5. Strike the wells sharply onto adsorbent paper or paper towels to remove all residual droplets.
6. Add 100 μ L of enzyme conjugate reagent into each well.
7. Incubate on an orbital micro-plate shaker at 100-150 rpm at room temperature (18-25°C) for 45 minutes.
8. Wash as detailed in 4 and 5 above.
9. Dispense 100 μ L of TMB Reagent into each well.
10. Gently mix on an orbital micro-plate shaker at 100-150 rpm at room temperature (18-25°C) for 20 minutes.
11. Stop the reaction by adding 100 μ L of Stop Solution to each well.
12. Gently mix. *It is important to make sure that all the blue color changes to yellow.*
13. Read the optical density at 450 nm with a microtiter plate reader within 15 minutes.

CALCULATION OF RESULTS

1. Calculate the average absorbance values (A_{450}) for each set of reference calibrators, and samples.
2. Construct a calibration curve by plotting the mean absorbance obtained from each reference calibrator against its concentration in ng/mL on linear graph paper, with absorbance values on the vertical or Y-axis and concentrations on the horizontal or X-axis.
3. Using the mean absorbance value for each sample, determine the corresponding concentration of α -1-AGP in ng/mL from the calibration curve.
4. Multiply the derived concentrations by the dilution factor to determine the actual concentration of α -1-AGP in the serum/plasma sample.
5. If available, PC graphing software may be used for the above steps.

6. If the OD₄₅₀ values of samples fall outside the calibration curve when tested at the suggested dilution of 100,000, samples should be diluted appropriately and re-tested.

TYPICAL CALIBRATION CURVE

A typical calibration curve with optical density readings at 450 nm on the Y axis against α -1-AGP concentrations on the X axis is shown below. This curve is for the purpose of illustration only and should not be used to calculate unknowns. Each user should obtain his or her data and calibration curve in each experiment.

α -1-AGP (ng/mL)	Absorbance (450 nm)
50	2.940
25	1.741
12.5	0.991
6.25	0.540
3.13	0.347
1.56	0.257

STORAGE

The unused kit should be stored at 4°C and the microtiter plate should be kept in a sealed bag with desiccant to minimize exposure to damp air. Test kits will remain stable until the expiration date provided that the components are stored as described above.

FOR RESEARCH USE ONLY

KAMIYA BIOMEDICAL COMPANY

12779 Gateway Drive, Seattle, WA 98168

Tel: (206) 575-8068 Fax: (206) 575-8094

Email: LifeScience@k-assay.com

www.k-assay.com